Determination of Iron(II), Iron(III) and Total Iron in Some β -Thalassemia Patients Using Different Analytical Techniques

Nabil Fakhre and Dashty Ali

Department of Chemistry, College of Education, University of Salahalddin, Erbil 00964, Iraq

Received: February 17, 2013 / Accepted: March 27, 2013 / Published: May 20, 2013.

Abstract: There are many well-known analytical methods for determination of iron(II) and iron(III). Among these methods: Gravimetric, titrimetric, potentiometric, conductometric and batch and flow-injection spectrophotometric methods. In present study, two batch spectrophotometric, atomic absorption spectrometric and biolabo kit methods have been used for determination of iron(II), iron(III) and total iron. The present methods have the advantages of high sensitivity, low cost reagent, low operation cost, simplicity, speed and their applications for determination of iron(II) and iron(III) in some serum samples of normal human and β -thalassemia patients in Erbil city. For the first time especially in Erbil city attempts were made to use zero, first and second derivative spectra to identify the serum samples of some β -thalassemia patients from the normal human serum samples due to the appearance and resolution of peaks in both cases.

Key words: Determination, iron, β -thalassemia patients, different analytical techniques.

1. Introduction

Iron as an essential element for living organisms. It has important role in oxygen transporter and oxidative-reductive reaction, and it can be utilized as alternative ferrous-ferric states [1]. The biological importance of iron is widely attributed to its chemical properties, there are two oxidative forms, ferrous (Fe^{2+}) and ferric (Fe^{3+}) iron, and iron is able to accept and donate electron. The living organisms should be able to control the internal amount of iron and to appropriately respond by changing its entry and storage processes. Failure in this control results in anemia or iron overloads [2]. Thalassemia is a pediatric inherited disease; it is a type of chronic, microcytic anemia disease [3]. It is gene-linked disease that can cause serious health problems because it can lead to the destruction of red blood cells [4].

Thalassemia is classified in two major sub-types alpha and beta thalassemia, which are results of mutation of the α -globin and β -globin genes, respectively [5]. α -thalassemia is classified to four different forms: one-gene deletion (silent carriers), two-gene deletion (α -thalassemia trait), HbH (three-gene deletion) disease and four-gene deletion (hydrops fetalis). β -thalassemia is classified to three different forms: β -thalassemia minor (or heterozygous), β -thalassemia intermedia (or β^+ -thalassemia) and β -thalassemia major (or β^0 -thalassemia). Many methods such as spectrophotometry, atomic absorption spectrometry, high performance liquid chromatography, polarography [6-8], ion selective electrode and high performance capillary electrophoresis have been proposed for determination of iron. This paper describes UV/visible spectrophotometric, atomic absorption spectrometric, derivative spectroscopy, biolabo kit methods and optimum conditions for the direct determination of iron(II) and iron(III) using

Corresponding author: Nabil Fakhre, Dr., Prof., main research field: environmental analytical chemistry. E-mail: havras@yahoo.com.

1,10-phenanthroline and thiocyanate reagents. The aim of this study was to use some techniques (e.g., UV/visible spectrophotometric methods, atomic absorption spectrometry, derivative spectroscopy and biolabo kit) for the determination of iron (II) and (III) ions and total iron in the blood serum samples of some of the β -thalassemia patients. The study also aims for distinguishing the serum samples of normal human and β -thalassemia patients

2. Materials and Methods

2.1 Apparatus

Spectral measurements were carried out on a CECIL CE 3021UV/visible digital spectrophotometer using 1 cm quartz cells A NV 203. Spectrophotometer was used for all absorbance measurements with a 1 cm glass cell. pH Meter 7020 (England) was used to measure pH values. SP 9 Atomic Absorption Spectrophotometer Philips with iron hollow cathode lamp at wavelength 248.3 nm was used.

2.2 Reagents

Chemicals used are of the highest purity available.

3. Results and Discussion

3.1 Simultaneous Determination of Iron(II) and Iron(III)

Synthetic solutions were prepared, containing 5.0 $\mu\text{g/mL}$ of iron(II) and different amounts of iron(III). It was found that iron(III) can be determined in the presence of 5.0 $\mu\text{g/mL}$ of iron(II) within the concentration range 2.0-10.0 $\mu\text{g/mL}$ iron(III).

3.2 Simultaneous Determination of Iron(III) and Iron(II)

Synthetic solutions were prepared, containing 5.0 $\mu\text{g/mL}$ of iron(III) and different amounts of iron(II). It was found that iron(II) can be determined in the concentration range of 3.0-5.5 $\mu\text{g/mL}$ depending upon the decreasing of the absorbance of

iron(III)—thiocyanate complex.

3.3 Atomic Absorption Method

The AAS (atomic absorption spectrometric) method is characterized by excellent selectivity and good detection limit under optimum condition and relative freedom from interferences [9].

The calibration curves obtained showed good linearity over the range 1.0-5.0 $\mu\text{g/mL}$ of iron. The molar absorptivity and detection limit were $0.11 \times 10^4 \text{ mol}^{-1} \cdot \text{cm}^{-1}$ and 0.90 $\mu\text{g/mL}$, respectively.

3.4 Zero and Derivative Spectroscopic Studies

The zero, first and second derivative spectra of some of the normal human and β -thalassemia patient serum samples were recorded. Figs. 1-4 show examples of the spectra of normal and patients serum samples, in which the normal human spectra show maximum absorptions at 417 nm and 280.5 nm, while the spectra of patient serum samples show the same maximum absorption at 417 nm with a shoulder after 450 nm. The sharp and smooth peak at 280.5 nm in the case of the spectra of normal human serum samples converted to multi peaks at the same region, in the case of β -thalassemia patients.

As expected, there are resolutions in the peaks and

Fig. 1 First derivative spectrum of diluted 1:10 normal human serum.

Fig. 2 First derivative spectrum of diluted 1:10 β -thalassemia human serum.

Fig. 3 Second derivative spectrum of diluted 1:10 normal human serum

Fig. 4 Second derivative spectrum of diluted 1:10 β -thalassemia human serum.

sharper peaks are obtained for the serum samples of β -thalassemia patients than the normal human. Therefore, either zero, first and second derivative spectra may be used for distinguishing the serum samples of normal human and β -thalassemia patients.

4. Conclusions

In the present study, two batch spectrophotometric, atomic absorption spectrometric and biolabo kit methods have been used for determination of iron(II), iron(III) and total iron. In addition zero, first and second derivative spectra were used to distinguish the some serum samples of normal human and β -thalassemia patients. In the present investigation, the total iron was determined in the mentioned serum samples using atomic absorption spectrometric method and biolabo kits. The results showed that atomic absorption spectrometry is more sensitive and accurate than the biolabo kits which are used in local laborites in Erbil city.

For the first time especially in Erbil city attempts were made to use zero, first and second spectra to identify the serum samples of some β -thalassemia patients from the normal human serum samples due to the appearance and resolution of peaks in both cases.

References

- [1] J.M. Sosa, I. Samarjija, L. Honovic, B. Jurisic, N-phthaloyl-glycine-hydroxamic acid as serum iron chelator in rats, *Acta. Pharm.* 58 (2008) 231-236.
- [2] A.M.D. Souza, Iron metabolism, *Einstein* 3 (2) (2005) 123-126.
- [3] R.N. Sawh, M. Martin, D. Baniewicz, Thalassemia, the facts and the controversies, *Catlin Anita Journal* 29 (6) (2003) 447-449.
- [4] P. Vang, O. Zongrum, R. Sindhuphak, N. Dusitsin, Preliminary study on thalassemia screening and genetic counseling in selective hmong people in Saraburi Province, *Hmong Studies Journal* 8 (2004) 1-19.
- [5] P. Charoenkwan, P. Thanarattanakorn, S. Chaovaluksakul, S. Sittipreechacharn, Rattika sae-tang and torpong sanguansermisri, hematological and molecular characterization of beta-thalassemia/Hb tak compound heterozygote, *Southeast Asian J. Trop Med. Public Health* 34 (2) (2003) 415.

- [6] K. Takeshi, I. Katsuhiko, Spectrophotometric determination of iron(II) by extraction with 2,2-dipyridylketoxime, Japan Analyst Journal 30 (40) (1980) 256-259.
- [7] M. Satake, H.B. Singh, Spectrophotometric determination of iron(II) after separation by adsorption of its N-benzoyl-N-phenyl-hydroxylamine complex on naphthalene, Def. Sciences Journal 31 (3) (1981) 241-244.
- [8] R. Escobar, J.M.C. Pavon, Selective spectrophotometric determination of trace amounts of iron with di(2-pyridyl)-NN-di [(8-quinolyl) amino] methane: Determination of iron in blood serum, Analyst 108 (1983) 821-826.
- [9] E. Pungor, A Practical Guide to Instrumental Analysis, CRC Press, Boca Raton, Florida, 1995.