Development of Immunosensor for Early Detection of Aquatic Harmful Algal Blooms (HABs)

Faridah Salam¹, Hazana Razali¹, Gayah Abdul Rahman¹, Roziawati Mohd Razali², Mohd Nor Azman Ayub², Syah Noor Muhamad Ramli¹ and Norhafniza Awaludin¹

1. Biotechnology and Nanotechnology Research Centre, Malaysian Agricultural Research and Development Institute (MARDI), MARDI Headquarters, Persiaran MARDI-UPM, Serdang 43400, Selangor, Malaysia

2. Fisheries Research Institute(FRI), Batu Maung 11960, Penang, Malaysia

Abstract: Blooms of toxic and harmful microalgae, known as “red tides”, represent a significant and tremendous threat to human health and fisheries resources throughout the Southeast Asia region and the world. A harmful microalga which produces secondary metabolite known as saxitoxin (STX) is a neurotoxin produced by a variety of genera of dinoflagellates. The ingestion of this toxin through contaminated shellfish will inhibit the neuron depolarizations and action potentials which result in respiratory failure in human and known as PSP (paralytic shellfish poisoning). PSP takes effect as soon as five minutes depending on the species and the concentration consumed. Currently, the identification of dinoflagellates cells is done by taxonomy, which is based on a broad expertise of specially trained staff, expensive equipment like electron microscopes and is very time consuming. In many countries, for the toxin monitoring, they use mouse bio-assay which is bulky, expensive, loss of animal life and high variability and higher limit of detection (40 µg per 100 g meat) and HPLC methods required many pre-treatment steps for the analysis. Therefore, in order to overcome this problem, development of early detection of harmful algae bloom and bio-toxin in aquatic ecosystem is needed. The detection is based on screen-printed carbon working electrode with onboard carbon counter and silver-silver chloride pseudo-reference electrode. A direct ELISA (enzyme-linked immunosorbent assay) format was developed and optimized on the surface of a carbon screen-printed working electrode by immobilizing the capture antibody using electro-deposition of gold nanoparticles conjugated with polyclonal anti-toxin antibodies. The detection reagent, toxin-HRP (horseradish peroxidase) conjugate was used as an enzyme label. Electrochemical detection was then carried out using 3,3',5,5'-tetramethylbenzidine dihydrochloride (TMB)/H₂O₂ as the enzyme mediator/substrate system and conducted using chronoamperometry at 100 mV vs. onboard screen-printed Ag-AgCl pseudo- reference electrode. The minimum detection limits for harmful micro algae about 1 cell per mL. The immunosensor is very selective which is able to detect *Alexandrium minutum* with 100% selectivity and minimal cross reaction with others nontoxic algae (< 2%).

Key words: Gold nanoparticles, immunosensor, screen-printed carbon electrode, *Alexandrium minutum*, saxitoxin, rapid detection, harmful algal blooms.

1. Introduction

HABs (harmful algal blooms) have become a significant threat to fisheries, aquaculture industries, tourism, public health and worldwide economies. In the case of long run effect on the massive blooms can cause human food poisoning by accumulating of algal-origin toxins in filter-feeding shellfish. The economic impact of harmful algae is increasing sharply than in the past due to the drastic increase of

consumption of seafood, growth of coastal inhabitants and tourism industries. A report published from USA indicated the massive impact of HABs on huge climate change and ocean acidification. The recurrence phenomenon of HABs is one of the big problems in Malaysia since it is now affecting Peninsular Malaysia. The demand for fish is expected to increase from 1.3 million tons in 2010 to 1.9 million metric tons in the year 2020 with a growth of 3.8% annually, therefore a great loss results in depletion of aquaculture and fishery productions in

Corresponding author: Faridah Salam, Ph.D., research field: biosensor technology.

long term period. This is a crucial issue for Malaysia as the aquaculture industry contributed RM 7.3 billion or 17.9% to the agro-food industry. Scientists have characterized 40 species of harmful microalgae in Malaysia and majority of these algae will threaten human and livestock. Currently, the identification of dinoflagellates cells is done by taxonomy, which is based on a broad expertise of specially trained staff, expensive equipment like electron microscopes and is very time-consuming. Due to this problem, aquaculture industries need a simple and real time for in-situ analysis of dinoflagellates producing saxitoxin detection.

Currently, biosensor devices can offer a very attractive alternative technology for contaminant detection since they can be rapid, sensitive and simple to perform [1] as well as can provide real-time and on-site analysis [2]. Real-time detection of chemical contaminants is important since it provides immediate interactive information regarding the sample being tested and enables food manufacturers to take corrective measures before the product is released for consumption. This method has been widely used in many applications such as in pollution control and monitoring of mining, industrial and toxic gases, environmental monitoring, quality control, drug development, agricultural and veterinary analysis and crime detection [2].

2. Material and Methods

2.1 Source of Chemicals

PBS (phosphate buffered saline), comprising of 0.13 mM NaH_2PO_4 , 0.5 mM Na_2HPO_4 and 0.51 mM NaCl, pH 7.4 was prepared by dissolving five buffer tablets in 1 L distilled-deionised water. For the analysis, ready-made TMB (3,3',5,5'-tetramethylbenzidine) solution was used. According to the manufacturer (Europa Bioproducts Ltd., Ely, UK), the TMB solution is stable at room temperature and is not sensitive to normal laboratory light. It is optimized with respect to TMB and

hydrogen peroxide concentrations and yields a linear response with the concentrations of HRP (horseradish peroxidase) usually employed in immunological assays. It also contains stabilizers.

2.2 Isolation and Propagation of Dinoflagellates *Alexandrium* spp.

The pure culture of the *Alexandrium minutum* was isolated from culture collection in Fisheries Research Institute. The bio-toxin produced algae, *Alexandrium minutum* was cultivated in 2-10 L flask (Fig. 1) with normal aeration for 2-6 weeks in the laboratory conducted by Fisheries Research Institute, Batu Maung, Pulau Pinang. After 6 weeks of cultivation, the *Alexandrium* cells were harvested by centrifugation (5,000 rpm, 10 min) and the pellet cells was collected and freeze-dried before the extraction process of bio-toxin and protein marker was proceed.

2.3 Dinoflagellates Bio-Toxin Purification

Bio-toxin purification was conducted following Cedric Robillot et al. [3] methods. The 5 g dried *Alexandrium* cells was grinded with 50 mL of 80% ethanol (pH 2 in HCL) ratio 1:1 (w/v) and was heated at 100 °C for 5 h under reflux to convert the majority of C-toxins to gonyautoxin. The extract was then applied to the activated charcoal and the toxin was eluted with 20% ethanol containing 6% acetic acid.

Fig. 1 *Alexandrium minutum* at 4-6 weeks culture in 10 L flask.

Eluted gonyau-toxins were further purified on a

Bio-Gel P2 column and then were eluted isocratically with 0.05% acetic acid. Toxic fractions were combined and concentrated by rotary evaporation under vacuum and lyophilized for conversion to STX. The toxin sample was re-dissolved in 0.05 M acetic acid and then was injected on a Bio-Rex 70 column and the toxin was eluted with a gradient of 0.05-3 M acetic acid with 10 mL fractions collected. Fig. 2 showed the chromatogram of the bio-toxin fraction after purification using the three columns (Charcoal, Bio-Gel P2 and Bio-Rex 70). The chromatogram of HPLC analysis of the peak that was eluted from the three columns indicated the type of bio-toxin (STX) obtained from the *Alexandrium minutum*.

2.4 Dinoflagellates Protein Biomarker Purification

The 5 g dried *Alexandrium* cells was grinded with 50 mL phosphate buffer saline (PBS, 0.01 M, pH 7.4) and centrifuged for 15 min at 6,000 rpm. The supernatant was collected and precipitated with 80% saturated ammonium sulphate for 30 min. This pellet was re-dissolved with PBS and was dialyzed three times in 5 L PBS (0.01 M, pH 7.4). *Alexandrium* cell wall protein from pellet (from the PBS extraction stage) was extracted again with 50 mL of 50 mM phosphate buffer containing 1 mM EDTA and 1 mM DTT by re-grinding, precipitated and centrifugation following the similar PBS extraction step. Extracted protein marker was further purified using Sephadex75 gel chromatography with 0.2 MPa and 0.2 mL·min⁻¹ of the flow rate using ACTA protein purifier system. Protein sample was then eluted by PBS and its content was

tested using BCA procedure for each of the fraction using BSA (bovine serum albumin) as standard protein.

2.5 Antibody Production against Protein Bio Marker and Bio-Toxin

The polyclonal antibodies were produced by immunizing New Zealand white rabbits via subcutaneous injection with an emulsion consisting of 0.5 mL of bio-toxin (5 µg·mL⁻¹) in 0.5 mL of PBS and an equal volume of CFA (complete Freund's adjuvant). The injections were repeated three times weekly after the initial injection, substituting IFA (incomplete Freund's adjuvant) for complete adjuvant. A booster injection was given one month after the initial injection and was repeated at monthly intervals thereafter. The rabbits were bled for antibody titer determinations two weeks after each boost. Antisera against bio-toxin was diluted with distilled water (1:10) and then precipitated with 80% saturated ammonium sulphate with continuous stirring to precipitate serum proteins. The serum mixture was centrifuged (Avanti J-26 XP, Beckman Coulter, Inc. USA) at 20,000 rpm for 15 minutes at 4 °C. The pellet was resuspended in PBS and dialyzed 3X in PBS to remove ammonium salt and then eluted through a protein A column (manually pack) using AKTA purifier instrument (Pharmacia Ltd., Sweden). Fractions giving the highest absorbance reading at OD 280 nm were collected and freeze-dried (IgG stock). As the freeze-drying did not affected the antibody activities and also will prolong the storage stability [4]. IgG titers for anti-bio-toxin antibody was

Fig. 2 Schematic diagram of the antibody-antigen reaction on the carbon working electrode using immunoassay format (a) Indirect competitive format (b) Sandwich format.

determined by indirect ELISA method.

2.6 Preparation of Conjugated Gold Nano-Particles with Antibody and Horseradish Peroxidase Enzyme

Gold nano-particles with average diameter of approximately 20 nm were purchased from BBInternational, Cardiff, UK. The colloidal gold solution was stored in a dark bottle at 4 °C and used directly without any pre-treatment. The particle size of colloidal gold was $\sim 20 \pm 3.2$ nm and the number of particles is $\sim 7 \times 10^{11}$ particles per mL. The UV-vis spectrometer showed an absorption peak at 525 nm. The antibody-colloid gold conjugate was prepared according to the procedure described by Chen et al. [5]. The antibody-colloidal gold conjugate was prepared by adding 100 μ L of anti-tetracycline antibody ($1.0 \text{ mg}\cdot\text{mL}^{-1}$) to 1.0 mL of pH-adjusted colloidal gold solution (pH 9.0), followed by slow shaking for one hour at room temperature. This allowed the antibody to adsorb onto the gold nanoparticles through a combination of ionic and hydrophobic interactions. Then, 100 μ L of 10% BSA solution was added to the gold-antibody mixture and left for further 30 minutes at room temperature. The mixture was then centrifuged at 9,000 rpm for 30 minutes. The supernatant solution was then discarded and the soft sediment of immuno-gold was dissolved in 100 μ L of 1% BSA solution and stored at 4 °C until used. The BSA was used to stabilize the immuno-gold colloid and minimize the non-specific adsorption during the assays and also to block the unoccupied site of the gold surface. The amount of antibody bound per nano-particle was then calculated based on protein assay results achieved from analysing the supernatant for unbound proteins using Bradford protein assay of Sigma, UK.

2.7 Screen-Printed Carbon Electrode Fabrication and Electrochemical Assays

SPCEs (screen-printed carbon electrodes) consisting of carbon working electrode, carbon

counter electrode and silver-silver chloride pseudo-reference electrode were fabricated by Screen Print Technology Sdn Bhd. Sungai Petani, Malaysia. The SPCE used in this work consisted of a carbon working electrode with a 5 mm diameter giving a 19.6 mm^2 planar area. All electrodes were then tested using a multimeter before use. The sensor edge connector was purchased from DropSense Ltd. UK through MetrohmSdnBhd, its product distributor is in Malaysia. Electrochemical measurements were carried out by placing a 50 μ L solution onto the electrode, covering the 3-electrode area. Each measurement was carried out in triplicates using a new strip in a non-deaerated and unstirred solution. Measurements were performed using the Autolab Type II (Eco Chemie, The Netherlands) with NOVA 1.6 software. Cyclic voltammetric measurements were carried out by scanning at 50 mV s^{-1} between -1.0 and 1.0 V relative to Ag/AgCl reference electrode. Stock solutions of 50 mM potassium ferrocyanide were prepared in 0.1 M KCl. For the selection of optimal potential for TMB- H_2O_2 -HRP system, chrono-amperometry was conducted similar to that reported previously [6].

2.8 Immunosensor Development

The carbon surface of the screen-printed electrode was directly used for the electro-deposition of polypyrrole-antibody-nano gold treatment. The freshly cleaned carbon was then covered with a 75 mM solution of polypyrrole plus antibody-nano gold and kept unstirred at room temperature. Chronoamperometry measurements were carried out by setting 1.0 V fix potential for 20 minutes onboard screen-printed Ag-AgCl pseudo-reference electrode. The antibody-nano gold coated electrode was then washed with 0.1 M phosphate buffer to remove the excess of unbound antibody-nano gold. The layer is very stable if kept dry (under silica gel) for several weeks at 4 °C [7]. For saxitoxin or Alexandrium assays, various dilutions of saxitoxin standard (0-250

ppb) or *Alexandrium* cells and saxitoxin-HRP conjugate or anti-*Alexandrium* antibody-HRP conjugate (20 μ L) were added to the electrodes and incubated for 20-30 minutes at 37 $^{\circ}$ C. The electrode was washed 3 times with PBS-T and dried gently. The assay was then performed by adding 50 μ L of TMB solution using chronoamperometry at 100 mV for 100-200 s. Calibration curve was fitted with a non-linear regression using 4-parameter logistic equation and the detection limit (LOD) was then calculated based on the following equation as described by Tijssen [8]:

$$\text{LOD} = x \left(\frac{a - d}{(a - d) - 3s} - 1 \right)^{-1/k}$$

where, s is for standard deviation of the zero value, a and d are the maximum and minimum values of calibration curve, x is the concentration at the EC50 value and k is the hill slope.

2.9 Cross Reactivity Studies

The sensitivity and specificity of the sensor was investigated in relation to other algae such *Pavlova*, *Teraselmis*, *VaunoBatan*, *Chaelo*, *Scipsiella*, *SK*, *Alexandrium affine* and *Alexandrium* spp. The specificity of this toxic *Alexandrium* (biotoxin) sensor with others nontoxic algae was conducted by replacing *Alexandrium minutum* with above listed algae.

3. Result and Discussion

The electrochemical immunosensor system developed in this work for saxitoxin and *Alexandrium minutum* detection was based on direct competitive and sandwich immunoassay format with HRP used as the enzyme label and TMB/H₂O₂ as the substrate/mediator system as illustrated in Fig. 2. The higher saxitoxin in the sample, the lower the signal achieved from the electrochemical immunosensor. While for the *Alexandrium minutum*, the higher cell numbers are in the sample, the higher signal is achieved in the electrochemical measurement.

3.1 Anti-*Alexandrium minutum* and Biotxin Antibody Characterization and Standard Curve Fabrication

The total protein from extracted *Alexandrium* from dried cells using ammonium sulphate precipitation is about 29.88% and after being further purified with Sephadex75 gel chromatography, the pure protein contain is about 8.45%. Protein profile from extracted sample was characterized and estimated about 115,000 KD. For the bio-toxin purification, HPLC and ELISA analysis is determined about 10.2% with 0.90 g STX from 5 g dried algae sample. Protein and bio-toxin complex were then used as an immunogen in immunization procedure for the antibody production. The primary and booster immunization in rabbit with bio-toxin complex with the addition of Freund's adjuvant is an inexpensive strategy for polyclonal antibody production. Freund's adjuvant (paraffin oil based) has been used for stimulation of the immune system by Mycobacterium in Complete Freund's Adjuvant to generate high antibody titers [9]. The initial experiment was to screen the purified polyclonal antibodies by titration with algae (*Alexandrium minutum*) as the target harmful bio-toxin. This assay involved the titration of monthly-purified antiserum samples using algae cells adsorbed to microtitre plates. Control experiments using purified pre immune serum showed insignificant binding levels. The whole pattern of antibodies level in rabbit within six months immunization schedule was obtained. There is a significant binding and a gradual increase in the apparent binding activity with boosting injections over the period tested. These initial tests provide information on the extent of serum maturation following the typical pattern on primary and secondary response, where specific antibodies appear after repeating immunizations. From the immunization schedule, bleed 3 and 4 showed the highest activity where the antibody titer of 1:0.00001 mg was obtained. The antibody concentration at 0.1 mg·mL⁻¹ was used for the development of algae-toxin standard curve using direct sandwich ELISA method.

The ELISA result for standard curve development shows a limit of detection about 5 cells mL⁻¹ that was obtained (Fig. 3a). Immunosensor technique was then developed using sandwich ELISA format by replacing anti-algae toxin complex antibody with nano-gold coated antibody on the sensor platform. Fig. 3b shows immunosensor standard curve that was developed using pure *A. minutum* cells. The LOD for *Alexandrium* detection is 1 cells mL⁻¹.

3.2 Specificity of the HABs Immunosensor

The sensitivity and specificity of the sensor was investigated in relation to other algae such *Pavlova*, *Teraselmis*, *VaunoBatan*, *Chaelo*, *Scipsiella*, *SK*, *Alexandrium affine* and *Alexandrium* spp. The specificity of this toxic *Alexandrium* (biotoxin) sensor

with others nontoxic algae was conducted by replacing *Alexandrium minutum* with above listed algae. The results showed that there is no cross-reactivity with non-toxic tested algae and the immunosensor is able to detect all the *Alexandrium* spp. tested with more than 50%.

3.3 Immunosensor Method Verification Using Actual Sample

Algae sample was collected from CherohPaloh, Pekan, Pahang. Nine samples out of 19 samples are positive *Alexandrium* measured by immunosensor strip and was correlated with microscopic analysis (Table 1). The data are the initial step of the application of HABs immunosensor for early diagnosis of HAB in aquatic ecosystem.

Fig. 3 Standard curve for *Alexandrium* using ELISA (a) and immunosensor (b).

Fig. 4 Cross-reactivity of immunosensor with others non-toxic algae.

Table 1 Comparative data of water samples analysis using immunosensor and microscopic analysis.

Sampling point	Name of sample	No. of <i>Alexandrium minutum</i> cells (cells/ml)				Microscope analysis	
		Rep 1	Rep 2	Mean	Mean * made to integer		
1	1.0 Plankton at 1m deep	11.58	8.07×10^{-9}	5.79000	6	No	
	1.1 Water sample at 1m	1.235	3.25×10^{-15}	0.61750	0	No	
	1.2 Water sample at 2m	0.104	0.014	0.05900	0	No	
	1.3 Water sample at 3m	96.52	1.02×10^{-8}	48.68675	48	Yes	
2	2.0 Plankton at 1m deep	3.567	6.73×10^{-8}	1.78350	2	No	
	2.1 Water sample at 1m	0.073	1.32×10^{-12}	0.05437	0	No	
	2.2 Water sample at 2m	0.04	2.16×10^{-5}	0.03063	0	No	
	2.3 Water sample at 3m	1.34×10^{-5}	5.25×10^{-6}	0.11575	0	No	
3	3.0 Plankton at 1m deep	0.015	0	0.00750	0	No	
	3.1 Water sample at 1m	9.153	0.117	4.63500	5	No	
	3.2 Water sample at 2m	9.99×10^{-5}	1.98×10^{-9}	0.00107	0	No	
	3.3 Water sample at 3m	1.235	0.541	1.23500	1	No	
4	4.0 Plankton at 1m deep	12.216	16.49	12.21600	12	Yes	
	4.1 Water sample at 1m	33.43	96.52	33.43000	33	Yes	
	4.2 Water sample at 2m	0.022	0.706	0.02200	0	No	
5	5.0 Plankton at 1m deep	0.771	3.17	0.77100	0	No	
	5.1 Water sample at 1m	4.514	1.374	4.51400	5	No	
	5.2 Water sample at 2m	10.179	6.428	10.17900	10	Yes	
	5.3 Water sample at 3m	0.211	4.906	0.21100	0	No	

4. Conclusion

One of the advantages of this immunosensor can be packed with customized portable digital reader and bio-reagent for easy use and handling. Another advantage of the present invention provides detection for the presence of toxic microalgae before HAB occurred and its toxin molecules in various fluids or solid food. A great advantage of direct detection of bio-toxin and toxic microalgae using antibody-nanogold-polymer electrode is direct sample application with minimal sample preparation/extraction. This is opposed to other bio-toxin and toxic microalgae determinations using established instrumentation (HPLC/LC-MS/Microscope) and mouse bioassay. In addition, the device components in the apparatus of the present invention are more robust than those used for example in optical measurements, facilitating the manufacture of the low cost reusable meter for use with disposable test devices. Further, the present invention can provide for very rapid test times.

Acknowledgements

The authors would like to thank the MOSTI for

funding this research project (02-03-08-SF0267/RB5018SF10).

References

- [1] Alocilja, E., and Radke, S. M. 2003. "Market Analysis of Biosensors for Food Safety." *Biosensors & Bioelectronics* 18: 841-6.
- [2] Tothill, I. E. 2011. "Biosensors and Nanomaterials and Their Application for Mycotoxin Determination." *World Mycotoxins Journal* 4: 361-74.
- [3] Cedric, R., Daniel, K., James, B., and Lyndon, E. L. 2009. "Synthesis of bifunctional saxitoxin analogues by biotinylation." *Toxicon* 53: 460-5.
- [4] Dráber, P., Dráberová, E., and Nováková, M. 1995. "Stability of Monoclonal Igm Antibodies Freeze-Dried in the Presence of Trehalose." *Journal of Immunological Methods* 181 (1): 37-43.
- [5] Chen, Z-P., Peng, Z-P., Zhang, P., Jin, X-F., Jiang, J-H., Zhang, X-B., Shen, G-L., and Yu, R-Q. 2007. *A Sensitive Immunosensor Using Colloidal Gold as Electrochemical Label Talanta* 72: 1800-4.
- [6] Salam, F., and Tothill, I. E. 2009. "Detection of Salmonella Typhimurium Using an Electrochemical Immunosensor." *Biosensors and Bioelectronics* 24: 2630-6.
- [7] Chaki, N., and Vijayamohan, K. 2002. "Self-assembled Monolayers as a Tunable Platform for Biosensor Applications." *Biosens. Bioelectron.* 17: 1-12.
- [8] Tijssen, P. 1985. *Practice and Theory of Immunoassay*.

Amsterdam: Elsevier.

- [9] Trott, D. L., Hellestad, E. M., Yang, M., and Cook, M. E.
2008. "Additions of Killed Whole Cell Bacteria

Preparations to Freund Complete Adjuvant Alter Laying
Hen Antibody Response to Soluble Protein Antigen."
Poultry Science 87: 912-7.